Caring for your

Bipolar and Monopolar Forceps

surgicalholdings

Don't use abrasive wire brushes to clean - they damage insulation

Connect bipolar or monopolar forcep to suitable test machine checking the insulation thoroughly

Check tips are clean with no Eschar

Ensure tips are not scratched or damaged

Check lead on set connects correctly with the forcep and safety test on suitable device

SMART

REPAIR OR REPLACE PROMISE

Key: Things to visually **check** Things to physically **test** General **information**

www.surgicalholdings.co.uk